



# Agilent 8890 Gas Chromatograph

## Safety Manual



## Notices

© Agilent Technologies, Inc. 2019

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

### Manual Part Number

G3540-90010

### Edition

Third edition, October 2019

Second edition, June 2019

First edition, January 2019

Printed in USA

Agilent Technologies, Inc.  
2850 Centerville Road  
Wilmington, DE 19808-1610 USA

安捷伦科技（上海）有限公司  
上海市浦东新区外高桥保税区  
英伦路 412 号  
联系电话：（800） 820 3278

### Warranty

**The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.**

### Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

### Restricted Rights Legend

U.S. Government Restricted Rights. Software and technical data rights granted to the federal government include only those rights customarily provided to end user customers. Agilent provides this customary commercial license in Software and technical data pursuant to FAR 12.211 (Technical Data) and 12.212 (Computer Software) and, for the Department of Defense, DFARS 252.227-7015 (Technical Data -Commercial Items) and DFARS 227.7202-3 (Rights in Commercial Computer Software or Computer Software Documentation).

### Safety Notices

#### CAUTION

A CAUTION notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

#### WARNING

A WARNING notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

# Introduction

Important Safety Warnings	4
Hydrogen Safety	7
Fume Hood	8
Electron Capture Detector (ECD)	9
Fuses and Batteries	10
Safety and Regulatory Certifications	11
Intended Use	15
Cleaning	15
Recycling the Product	15

## Important Safety Warnings

Before moving on, there are several important safety notices that you should always keep in mind when using the Agilent Agilent 8890.

### WARNING

**When handling/using chemicals for preparation or use within the GC, all applicable local and national laboratory safety practices must be followed. This would include, but is not limited to, correct use of Personal Protective Equipment (PPE), correct use of storage vials, using vent hoods, and correct handling of chemicals, as defined in the laboratory's internal safety analysis and standard operating procedures. Failure to adhere to laboratory safety practices could lead to injury or death.**

---

## Many internal parts of the GC carry dangerous voltages

If the GC is connected to a power source, even if the power switch is off, potentially dangerous voltages exist on:

- The wiring between the GC power cord and the AC power supply, the AC power supply itself, and the wiring from the AC power supply to the power switch.

With the power switch on, potentially dangerous voltages also exist on:

- All electronics boards in the instrument.
- The internal wires and cables connected to these boards.
- The wires for the oven heater.

### WARNING

**All these parts are shielded by covers. With the covers in place, it should be difficult to accidentally make contact with dangerous voltages. Unless specifically instructed to, never remove a cover unless the instrument is unplugged.**

---

### WARNING


**If the power cord insulation is frayed or worn, the cord must be replaced. Contact your Agilent service representative.**

---

## Lifting

### WARNING

The GC is heavy. To avoid injury, use a two-person lift.



### CAUTION

Make sure that all samplers, trays, and detectors have been removed prior to moving or relocating the GC. Prior to moving, all covers and enclosures must also be seated properly. Make sure all covers and enclosures are secure.

### CAUTION

Make sure your GC has adequate distance from the wall and other instruments.

## Do not use an Uninterruptable Power Supply (UPS) with a GC

If the area where the GC is located suddenly loses power, an unsafe condition can result if the GC remains powered on. Do not use the GC with a UPS.

## Electrostatic discharge is a threat to GC electronics

The printed circuit (PC) boards in the GC can be damaged by electrostatic discharge. Do not touch any of the boards unless it is absolutely necessary. If you must handle them, wear a grounded wrist strap and take other antistatic precautions. Wear a grounded wrist strap any time you must remove the GC right side cover.

## Many parts are dangerously hot

Many parts of the GC operate at temperatures high enough to cause serious burns. These parts include but are not limited to:

- The inlet area
- The oven and its contents
- The detector area
- The column nuts attaching the column to an inlet or detector
- The valve box

You should always cool these areas of the GC to room temperature before working on them. They will cool faster if you first set the temperature of the heated zone to room temperature. Turn the zone off after it has reached the setpoint. If you must perform maintenance on hot parts, use a wrench and wear thermally protective gloves.

Whenever possible, cool the instrument part that you will be maintaining before you begin working on it. Do not open the oven door immediately after the heating sequence.

### **WARNING**

---

**Be careful when working behind the instrument. During cool-down cycles, the GC emits hot exhaust which can cause burns.**

### **WARNING**

---

**The insulation around the inlets, detectors, valve box, and the insulation cups is made of refractory ceramic fibers. To avoid inhaling fiber particles, we recommend the following safety procedures: ventilate your work area; wear long sleeves, gloves, safety glasses, and a disposable dust/mist respirator; dispose of insulation in a sealed plastic bag; wash your hands with mild soap and cold water after handling the insulation.**

## Oven thermal leaks

### **WARNING**

---

**Objects passing through the oven door seal can cause thermal leaks which create hazardous hot spots which cause burns and melt equipment.**

Do not allow wiring or temperature probes to pass through the oven door jam. Agilent recommends using one of the access holes. If the access hole is used, replace the insulation so there are no thermal leaks.

# Hydrogen Safety

Hydrogen gas may be used as carrier gas, and/or as fuel for the FID, FPD, NPD, NCD, and SCD. When mixed with air, hydrogen can form explosive mixtures.

## WARNING

**When using hydrogen (H<sub>2</sub>) as the carrier gas or fuel gas, be aware that hydrogen gas can flow into the GC oven and create an explosion hazard. Therefore, be sure that the supply is turned off until all connections are made and ensure that the inlet and detector column fittings are either connected to a column or capped at all times when hydrogen gas is supplied to the instrument.**

**Hydrogen is flammable. Leaks, when confined in an enclosed space, may create a fire hazard. In any application using hydrogen, leak test all connections, lines, and valves before operating the instrument. Always turn off the hydrogen supply at its source before working on the instrument.**

---

Hydrogen is a commonly used GC carrier gas. Hydrogen is potentially explosive and has other dangerous characteristics.

- Hydrogen is combustible over a wide range of concentrations. At atmospheric pressure, hydrogen is combustible at concentrations from 4% to 74.2% by volume.
- Hydrogen has the highest burning velocity of any gas.
- Hydrogen has a very low ignition energy.
- Hydrogen that is allowed to expand rapidly from high pressure into the atmosphere can self-ignite due to an electrostatic spark.
- Hydrogen burns with a nonluminous flame which can be invisible under bright light.

## Hydrogen shutdown

Hydrogen gas may be used as a carrier or as fuel for some detectors. When mixed with air, hydrogen can form explosive mixtures.

The GC monitors inlet, detector, and auxiliary gas streams. If a stream shuts down because it is unable to reach its flow or pressure setpoint *and* if that stream is configured to use hydrogen, the GC assumes that a leak has occurred and declares a *hydrogen safety shutdown*. The effects are:

- The offending channel and any associated channels (such as septum purge) are set off.
- The split valves in the inlets open.
- The oven (heater and fan) turns off.
- The small heated zones are turned off.
- An alarm tone sounds.

To recover from this state, fix the cause of the shutdown (tank valve closed, serious leak, others). Turn the instrument off, then back on.

## 1 Introduction

### Measuring hydrogen gas flows

#### **WARNING**

The GC cannot always detect small leaks in inlet and/or detector gas streams. For this reason, it is vital that column fittings should always be either connected to a column, or have a cap or plug installed. The H<sub>2</sub> streams must be configured for hydrogen so that the GC is aware of hydrogen use.

---

### Measuring hydrogen gas flows

#### **WARNING**

Do not measure hydrogen together with air or oxygen. This can create explosive mixtures that can be ignited by the automatic ignitor.

To avoid this hazard turn the automatic ignitor off before you begin, and always measure gases separately.

---

## Fume Hood

During normal operation of the GC with many detectors and inlets, some of the carrier gas and sample vents outside the instrument through the split vent, septum purge vent, and detector exhaust. If any sample components are toxic or noxious, or if hydrogen is used as the carrier gas, these exhausts must be vented to a fume hood. Place the GC in the hood or attach a large diameter venting tube to the outlet for proper ventilation.



## Electron Capture Detector (ECD)

The ECD contains a cell plated with  $^{63}\text{Ni}$ , a radioactive isotope. The beta particles released at the energy level in the detector have little penetrating power—the surface layer of the skin or a few sheets of paper will stop most of them—but they may be hazardous if the isotope is ingested or inhaled. For this reason, handle the cell with care. Cap the detector inlet and outlet fittings when the detector is not in use. Never introduce corrosive chemicals into the detector. Vent detector exhaust outside the laboratory environment.

Refer to the safety documentation provided with the detector for important details about safety, maintenance, and compliance with local government regulation.

### WARNING

**Materials that may react with the  $^{63}\text{Ni}$  source, either to form volatile products or to cause physical degradation of the plated film, must be avoided. These materials include oxidizing compounds, acids, wet halogens, wet nitric acid, ammonium hydroxide, hydrogen sulfide, PCPs, and carbon monoxide. This list is not exhaustive but indicates the kinds of compounds that may cause damage to  $^{63}\text{Ni}$  detectors.**

### WARNING

**In the extremely unlikely event that the oven or the detector-heated zone should go into thermal runaway (maximum, uncontrolled heating in excess of 450 °C) and the detector remains exposed to this condition for more than 12 hours, take the following steps:**

- 1 After turning off the main power and allowing the instrument to cool to room temperature, cap the detector inlet and exhaust vent openings. Wear disposable plastic gloves and observe normal laboratory safety precautions.
- 2 Contact your local Agilent Technologies sales office or distributor for ECD disposal instructions
- 3 Include a letter stating the condition of abuse.

**It is unlikely, even in this very unusual situation, that radioactive material will escape the cell. However, permanent damage to the  $^{63}\text{Ni}$  plating within the cell is possible; therefore, the cell must be returned for exchange.**

### WARNING

**Do not use solvents to clean the ECD.**

### WARNING

**You may not open the ECD cell unless authorized to do so by your local nuclear regulatory agency. Do not disturb the four socket-head bolts. These hold the cell halves together. United States customers: removing or disturbing the four socket-head bolts is a violation of the terms of the exemption and could create a safety hazard.**

When handling ECDs:

- Never eat, drink, or smoke.
- Always wear safety glasses when working with or near open ECDs.
- Wear protective clothing such as laboratory jackets, safety glasses, and gloves, and follow good laboratory practices. Wash hands thoroughly with a mild nonabrasive cleaner after handling ECDs.

- Cap the inlet and outlet fittings when the ECD is not in use.
- Connect the ECD exhaust vent to a fume hood or vent it to the outside. See the latest revision of 10 CFR Part 20 (including Appendix B), or the applicable state regulation. For other countries, consult with the appropriate agency for equivalent requirements.

Agilent Technologies recommends a vent line internal diameter of 6 mm (1/4-inch) or greater. With a line of this diameter, the length is not critical.

## Fuses and Batteries

The GC requires two fuses (F1, F2) for proper operation. The replacement fuses are F1/F2-Ceramic body, 20 A, 250 V, Type F, but must only be accessed by Agilent trained service personnel.

There is a Real Time Clock component on the Logic board that includes a battery inside.

### **WARNING**

**Disconnect the appliance inlet coupler from the main supply before replacing the fuses.**

## Disconnect Device

### **WARNING**

**Appliance inlet coupler (main input power cord) is the power disconnect device. Do not position the instrument such that access to the coupler or plug is impaired.**

### **WARNING**


**The front power switch does not disconnect power for the instrument. Pressing the disconnect switch places the instrument in a standby mode. To avoid an electric shock hazard, disconnect the device by unplugging the power cord before performing instrument service.**

# Safety and Regulatory Certifications

The Agilent 8890 GC conforms to the following safety standards:

- Canadian Standards Association (CSA): C22.2 No. 61010-1
- CSA/Nationally Recognized Test Laboratory (NRTL): ANSI/UL 61010-1
- International Electrotechnical Commission (IEC): 61010-1, 61010-2-010, 61010-2-081
- EuroNorm (EN): 61010-1

The Agilent 8890 GC conforms to the following regulations on Electromagnetic Compatibility (EMC) and Radio Frequency Interference (RFI):

- CISPR 11/EN 55011: Group 1, Class A
- IEC/EN 61326-1
-  AUS/NZ

The Agilent 8890 GC complies with the following company standard:

- Q31/0115000033CXXX-2018
- The China Pattern certificate number is 2018CXXX-31.

This ISM device complies with Canadian ICES-001. Cet appareil ISM est conforme a la norme NMB-001 du Canada.

- CAN ICES-1/NMB-1, ISM 1-A



The Agilent GC is designed and manufactured under a quality system registered to ISO 9001. Declaration of Conformity available.

This product complies with the EU RoHS Directive 2011/65/EU and conforms to EN 50581.

## Regulatory compliance identification number

For the purpose of regulatory compliance certifications and identification, your product has been assigned a unique Agilent Regulatory Model Number (RMN). The 8890 GC RMN3540A can be found on the product nameplate label, along with all required approval markings and information. When requesting compliance information for this product, always refer to this RMN. The RMN should not be confused with the marketing name or model number of the product.

## EMC Declaration for South Korea

### 사용자안내문

이 기기는 업무용 환경에서 사용할 목적으로 적합성평가를 받은 기기로서 가정용 환경에서 사용하는 경우 전파간섭의 우려가 있습니다.

This equipment has been evaluated for its suitability for use in a commercial environment. When used in a domestic environment, there is a risk of radio interference.

## Detachable Power Cord Declaration for Japan

(弊社提供の電源コードセットが汎用性がない旨を示す)

### 電源コードセットの取扱いについて (日本国内向け)

**製品には、同梱された電源コードセットをお使いください。同梱された電源コードセットは、他の製品では使用できません。**

Notice - The power cords for Japanese market

Your product must only use the power cord that was shipped with this product. Do not use this power cord with any other product.

## Information

The Agilent Gas Chromatograph meets the following IEC (International Electro-technical Commission) classifications: Safety Class 1, Transient Overvoltage Category II, Pollution Degree 2.

This unit has been designed and tested in accordance with recognized safety standards and is designed for use indoors in non-classified locations. If the instrument is used in a manner not specified by the manufacturer, the protection provided by the instrument may be impaired. Whenever the safety protection of the Agilent Gas Chromatograph has been compromised, disconnect the unit from all power sources and secure the unit against unintended operation.

Refer servicing to qualified service personnel. Substituting parts or performing any unauthorized modification to the instrument may result in a safety hazard.

## Symbols

The apparatus is marked with this symbol when the user should refer to the instruction manual in order to protect risk of harm to the operator and to protect the apparatus against damage.



Indicates a hot surface.



Indicates hazardous voltages.



Indicates earth (ground) terminal.



Indicates potential explosion hazard.



Indicates radioactivity hazard.



Indicates electrostatic discharge hazard.



Indicates a hazard. See the Agilent GC user documentation for the item labeled.



Indicates that you must not discard this electrical/electronic product in domestic household waste



To prevent personal injury, two person lift recommended



Manufacturing date.



Power symbol indicates **On/Off**. The apparatus is not completely disconnected from the main supply when the power switch is in the **Off** position.



Warnings in the manual or on the instrument must be observed during all phases of operation, service, and repair of this instrument. Failure to comply with these precautions violates safety standards of design and the intended use of the instrument. Agilent Technologies assumes no liability for the customer's failure to comply with these requirements.

## Electromagnetic compatibility

This device complies with the requirements of CISPR 11. Operation is subject to the following two conditions:

- This device may not cause harmful interference.
- This device must accept any interference received, including interference that may cause undesired operation.

If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try one or more of the following measures:

- 1 Relocate the radio or antenna.
- 2 Move the device away from the radio or television.
- 3 Plug the device into a different electrical outlet, so that the device and the radio or television are on separate electrical circuits.
- 4 Make sure that all peripheral devices are also certified.
- 5 Make sure that appropriate cables are used to connect the device to peripheral equipment.
- 6 Consult your equipment dealer, Agilent Technologies, or an experienced technician for assistance.
- 7 Changes or modifications not expressly approved by Agilent Technologies could void the user's authority to operate the equipment.

## Sound Emission Certification for Federal Republic of Germany

### **Sound pressure**

Sound pressure  $L_p < 70$  dB(A) according to DIN-EN 27779.

### **Schalldruckpegel**

Schalldruckpegel  $LP < 70$  dB(A) nach DIN-EN 27779.

## Intended Use

Agilent products must only be used in the manner described in the Agilent product user guides. Any other use may result in damage to the product or personal injury. Agilent is not responsible for any damages caused, in whole or in part, by improper use of the products, unauthorized alterations, adjustments or modifications to the products, failure to comply with procedures in Agilent product user guides, or use of the products in violation of applicable laws, rules or regulations.

## Cleaning

To clean the unit, disconnect the power and wipe down with a damp, lint-free cloth.

## Recycling the Product



Instructions for Disposal of Waste Equipment by Users in the European Union. This symbol on the product or its packaging indicates that this product must not be disposed of with other waste. Instead, it is your responsibility to dispose of your waste equipment by handing it over to a designated collection point for the recycling of waste electrical and electronic equipment. The separate collection and recycling of your waste equipment at the time of disposal will help conserve natural resources and ensure that it is recycled in a manner that protects human health and the environment. For more information about where you can drop off your waste equipment for recycling, please contact your local city recycling office or the dealer from whom you originally purchased the product.

For recycling, contact your local Agilent sales office.

[www.agilent.com](http://www.agilent.com)

© Agilent Technologies, Inc. 2019

Third edition, October 2019



G3540-90010

