


Calibration of Limonene in a California Lemon Oil Sample by GC-TOFMS

Life Science and Chemical Analysis Centre
LECO Corporation, Saint Joseph, Michigan USA



A calibration curve for limonene in lemon oil for concentrations from 8.5 pg/ μ l to 850 pg/ μ l.

Column:

10 m x 0.18 mm x 0.2 μ m Rtx-5

TOFMS:

35 to 350 m/z @ 10 spectra/s

For questions on this analysis e-mail us at:
life_science@leco.com

For a complete listing of snapshots and application notes visit us on the web at www.leco.com

